

AI-enabled POCUS for breast cancer risk stratification in a resource-limited tertiary clinic

**Authors:**

Kathryn Malherbe¹ 
Francois Malherbe² 
Liana Roodt³ 

Affiliations:

¹Department of Imaging,
Faculty of Health Sciences,
Malherbe Imaging Inc,
Pretoria, South Africa

²Department of Surgery,
Faculty of Health Sciences,
University of Cape Town,
Cape Town, South Africa

³Division of General Surgery,
Department of Surgery,
Faculty of Health Sciences,
University of Cape Town,
Cape Town, South Africa

Corresponding author:

Kathryn Malherbe,
kathryn@medsol.ai

Dates:

Received: 07 May 2025

Accepted: 01 Aug. 2025

Published: 09 Oct. 2025

How to cite this article:

Malherbe K, Malherbe F,
Roodt L. AI-enabled POCUS
for breast cancer risk
stratification in a resource-
limited tertiary clinic. *S Afr J
Rad.* 2025;29(1), a3195.
<https://doi.org/10.4102/sajr.v29i1.3195>

Copyright:

© 2025. The Authors.
Licensee: AOSIS. This
work is licensed under
the Creative Commons
Attribution 4.0 International
(CC BY 4.0) license (<https://creativecommons.org/licenses/by/4.0/>).

Read online:

Scan this QR
code with your
smart phone or
mobile device
to read online.

Background: Breast cancer remains a major public health burden in South Africa, with diagnostic delays contributing to poor outcomes. Ultrasound is effective for early detection but is limited by access and operator variability. Integrating artificial intelligence (AI) into point-of-care ultrasound (POCUS) offers a potential solution.

Objectives: To evaluate the diagnostic performance of a locally developed AI-enabled POCUS system (Breast AI) in predicting malignancy among women with palpable breast abnormalities.

Method: A prospective cohort study was conducted between June 2024 and November 2024 at Groote Schuur Hospital. Women aged ≥ 25 years with suspicious breast lesions underwent Breast AI ultrasound prior to biopsy. Real-time malignancy risk scores were compared with histopathological results. Diagnostic accuracy was assessed using sensitivity, specificity, positive predictive value (PPV), F1 score and area under the curve (AUC).

Results: Among 159 participants, Breast AI achieved a sensitivity of 67.2%, specificity of 79.4% and PPV of 70.3% at a 51% threshold. The AUC was 0.76, reflecting moderate discriminatory performance. F1 score analysis identified 51% as the optimal cut-off (F1 = 65.7%). Benign pathologies such as fibroadenomas and fat necrosis correlated with low AI scores. A three-tiered risk model was developed: $< 30\%$ (low), $30\% - 51\%$ (intermediate) and $> 51\%$ (high risk).

Conclusion: Breast AI demonstrates promising diagnostic accuracy for triaging suspicious breast lesions, particularly in resource-constrained settings.

Contribution: This study provides real-world evidence supporting the integration of AI into POCUS to improve breast cancer detection and clinical decision-making in low-resource environments.

Keywords: Breast AI; point-of-care ultrasound; breast cancer; diagnostic triage; artificial intelligence.

Introduction

Breast cancer remains a major public health burden in South Africa, with women facing a 1 in 29 lifetime risk of developing the disease.¹ Despite increasing incidence rates, delays in diagnosis and assessment, exacerbated by systemic inefficiencies and sociocultural barriers, contribute significantly to poor clinical outcomes.^{2,3}

In response, South Africa's Breast Cancer Prevention and Control Policy emphasises clinical breast examinations (CBEs) and breast self-examinations (BSEs) as essential early detection strategies, particularly in the absence of a national mammographic screening programme.⁴ While these modalities are endorsed by the National Department of Health, implementation has been hindered by limited resources, inconsistent training, and low levels of public awareness.³ Multimedia educational campaigns have sought to improve community engagement and encourage early presentation.⁵

Ultrasound is a critical contributory imaging modality in breast cancer detection,^{6,7,8} particularly for women with dense breast tissue or in settings where mammography access is limited. It enhances the characterisation of clinically palpable malignancies and is widely recognised as an effective adjunct to CBEs.^{6,7} Within South Africa, increasing the availability of ultrasound – especially point-of-care ultrasound (POCUS) – has been identified as a strategic priority.⁸ Conventional ultrasound typically involves a cart-based system operated in radiology suites, whereas POCUS refers to

portable or handheld ultrasound devices used directly at the patient's bedside or in primary care settings, enabling real-time clinical decision-making. However, broader integration of both modalities is constrained by disparities in equipment availability, inadequate practitioner training and workflow inefficiencies.^{9,10}

The introduction of artificial intelligence (AI) to breast ultrasound holds significant promise in addressing these limitations. Artificial intelligence algorithms have demonstrated high accuracy in lesion detection, reducing observer variability and improving diagnostic efficiency.^{11,12} In South Africa, an AI-based breast ultrasound system has recently received approval from the South African Health Products Regulatory Authority (SAHPRA) as a Class B medical device. This system has been trained on more than 40 000 histologically confirmed cases and demonstrated a reported diagnostic accuracy of 97.6%.¹³ The Breast AI application, integrated with a wireless Clarius™ handheld ultrasound probe and accessible via Android, enables real-time malignancy risk prediction.¹³

This study evaluated the performance of the Breast AI system in a high-volume tertiary care setting, assessing its diagnostic accuracy and utility in predicting breast malignancies in patients presenting with palpable abnormalities. Breast AI was employed to assess malignancy risk, with predictions cross-referenced against histopathological findings. The model's performance was evaluated using standard diagnostic metrics, namely sensitivity (the ability of Breast AI to correctly identify malignant cases), specificity (the ability of Breast AI to correctly classify benign cases) and positive predictive value (PPV) (the likelihood that a lesion flagged as high-risk by Breast AI corresponds to malignancy).

Research methods and design

Study design, setting and data collection

A prospective cohort study was conducted at the Groote Schuur Hospital diagnostic breast clinic from 01 June 2024 to 30 November 2024. A total of 174 participants were enrolled in this study. Fifteen patients were excluded because of the non-diagnostic nature of the biopsy specimen. Patients aged 25 years and older who presented to the breast clinic with a suspicious lesion (a palpable abnormality) requiring a biopsy, were included in the study. All biopsies were performed under ultrasound guidance. Patients excluded were those under 25 years of age, those with locally advanced tumours involving the skin, obvious inflammatory conditions (mastitis and breast abscess), and solitary axillary lymph nodes. Other exclusion criteria included patients who had a prior confirmed diagnosis of breast cancer before undergoing Breast AI analysis, had incomplete imaging data, or had no confirmed histological diagnosis after biopsy and were lost to follow-up before a final diagnosis could be determined.

The sample size was calculated using the formula for estimating a single population proportion, where Z is the

Z -score corresponding to a 95% confidence level (1.96), p is the expected proportion of malignancy (assumed at 0.5 for maximum variability), and d is the margin of error (set at 0.075 or 7.5%). To account for potential exclusions and data loss, an additional 15% was added, yielding a target sample size of approximately 197. Ultimately, 174 participants were enrolled, which retained adequate statistical power for the primary diagnostic performance analyses.

All participants underwent CBE followed by imaging with the Breast AI system, which integrates real-time malignancy risk scoring using a wireless Clarius™ ultrasound probe. The AI system provided percentage-based malignancy predictions at the point of care. All imaging and diagnostic data were prospectively recorded into a secure REDCap database. Collected variables included: patient demographics (age, gender), body mass index (BMI), presenting symptoms and their duration, breast volume (estimated on ultrasound), lesion location, AI prediction percentage, biopsy method, and final histology report.

The Breast AI system uses an artificial neural network (ANN)-based deep learning model that provides real-time probabilistic classification of malignancy risk. The model was trained on over 40 000 histologically confirmed ultrasound images and is capable of feature extraction from greyscale images, enabling contextual lesion interpretation. The system differs from traditional AI tools by integrating risk prediction into a handheld point-of-care device, promoting bedside utility. Three qualified operators participated: one radiologist and two mammographers, each with 10–15 years of breast imaging experience. This POCUS-based protocol, while distinct from conventional cart-based breast ultrasound systems, ensured portable and consistent acquisition suitable for triage applications.

Ultrasound image acquisition followed standard breast scanning protocols in alignment with the American College of Radiology (ACR) practice guidelines. All examinations were conducted using a Clarius L15 HD3 handheld linear array probe (5 MHz – 15 MHz), optimised for high-resolution superficial imaging. Operators were instructed to perform systematic radial and anti-radial scans across the lesion, including full coverage in the longitudinal and transverse planes. Scan depth was adjusted to include the chest wall, and gain, and focal zones were optimised in real-time to ensure uniform echogenicity. Images were acquired with minimal probe compression to preserve lesion architecture. Each breast was scanned in a clockwise manner from the 12 o'clock position, with annotations for clock-face location, depth, size and proximity to anatomical landmarks (e.g. nipple, Cooper's ligaments). Operators also evaluated lesion margins, echotexture and posterior acoustic features according to BI-RADS descriptors. All sonographers received standardised refresher training prior to the study to ensure consistency across image acquisition.

Data analysis

Descriptive statistics were computed using frequencies, proportions, medians, and interquartile ranges (IQRs) where appropriate. Continuous variables were compared using Student's *t*-test or one-way analysis of variance (ANOVA) for normally distributed data, and non-parametric equivalents (e.g. Mann–Whitney *U* or Kruskal–Wallis tests) for skewed distributions. To evaluate classification performance, confusion matrices and receiver operating characteristic (ROC) curves were constructed, and the area under the curve (AUC) was calculated. Raincloud plots were used to visualise the distribution of AI risk scores stratified by histological outcome. All statistical analyses were performed using R software version 4.3.2 (R Foundation for Statistical Computing, Vienna, Austria), and visualisations were generated using the ggplot2 and ggdist packages.

Additionally, a Breast AI-RADS (Artificial Intelligence – Reporting and Data System)–like categorisation scheme was employed to assess risk stratification thresholds. This system mirrors the BI-RADS framework, assigning Breast AI outputs, thereby standardising interpretation and supporting clinical decision-making. A confusion matrix was generated using a predefined risk threshold of 51%, identified via Youden's Index from the ROC analysis. Lesions with AI-generated malignancy risk scores $\geq 51\%$ were classified as 'high-risk', while those $< 51\%$ were deemed 'low-risk'. This structured approach enabled a comprehensive evaluation of Breast AI's ability to function as a diagnostic support tool, with findings compared to standard histological diagnoses. All analyses were performed using a significance threshold of $p < 0.05$.

Ethical considerations

Ethical approval was obtained from the University of Cape Town Health Research Ethics Committee (reference number 371/2024), and informed consent was acquired from all participants in advance. All data were entered into a password-protected REDCap database and exported anonymously to an Excel spreadsheet for analysis.

Results

A total of 174 participants were enrolled in this study. Fifteen patients were excluded because of the non-diagnostic nature of the biopsy specimen. The median age of the cohort was 48 years (IQR 38–60), with a median BMI of 29.5 kg/m² (IQR 25–34), placing the group in the overweight category. The median breast volume was 494 cc (IQR 325.5–699.5).

With the risk prediction threshold suggested at 51% for the Breast AI system, it correctly identified 45 malignant cases (true positives) and 73 benign cases (true negatives). It also incorrectly flagged 19 benign lesions as high-risk (false positives) and failed to detect 22 malignant cases (false negatives). These values were used to derive the system's core diagnostic performance metrics. The Breast AI system demonstrated moderate performance when benchmarked against clinical standards (Table 1).

The F1 score serves as a balanced measure of diagnostic performance, particularly when both false positives and false negatives carry clinical consequences. It is defined in Equation 1 as:

$$F_1 = 2 \times (\text{Precision} \times \text{Recall}) / (\text{Precision} + \text{Recall}) \quad [\text{Eqn 1}]$$

$$F_1 = 2 \times (70.3 \times 67.2) / (70.3 + 67.2) \approx 68.7\%$$

$$\text{Precision} = TP / (TP + FP) = 45 / (45 + 19) \approx 70.3\%$$

$$\text{Recall} = TP / (TP + FN) = 45 / (45 + 22) \approx 67.2\%$$

Further analysis of diagnostic performance by lesion laterality revealed asymmetry in system performance, with AI sensitivity and specificity higher for right-sided lesions. Table 2 highlights the performance discrepancies between left and right breast lesion detection.

The ROC curve for the Breast AI system yielded an AUC of 0.76, indicating moderate to strong discriminatory power between benign and malignant lesions (Figure 1). The optimal malignancy risk threshold of 51%, determined by Youden's Index, was used for binary classification throughout this study (Figure 1).

The distribution of AI-generated malignancy risk scores is illustrated in the raincloud plot (Figure 2). The results demonstrated a bimodal distribution, with most benign lesions clustering below 30% and malignant lesions


TABLE 1: Comparative diagnostic performance metrics.

Metric	Breast AI (%)	Diagnostic benchmark (%)
Sensitivity	67.2	85.0
Specificity	79.4	80.0
Positive predictive value	70.1	70.0

AI, artificial intelligence.

TABLE 2: Diagnostic performance by lesion laterality.

Side	True positives	False negatives	True negatives	False positives	Sensitivity (%)	Specificity (%)
Right	12	2	15	2	85.7	88.2
Left	10	7	17	3	58.8	85.0



Note: The red 'O' indicates the optimal cut-off point of 51% based on Youden's Index. ROC, Receiver operating characteristic.

FIGURE 1: Receiver operating characteristic (ROC) curve for breast artificial intelligence system: ROC curve (area under the curve: 0.768).

predominantly distributed above 51%. A diagnostic 'grey zone' was identified between 30% and 51%, where score overlap created potential classification ambiguity.

This suggests the potential clinical benefit of a three-tiered risk stratification system, categorising cases into low-risk (< 30%), intermediate-risk (30%–50%) and high-risk ($\geq 51\%$) categories to support more nuanced triage decisions. This is seen in Figure 3:

The benign lesions were classified into four groups based on final histology: (1) abscesses or inflammatory conditions and fibrosis, (2) fibroadenomas or other fibroepithelial lesions, (3) fat necrosis and (4) all remaining benign entities. ANOVA revealed no statistically significant differences among the groups ($p=0.79$). Corresponding box plots are presented in Figure 4. Further breakdown of the benign entities termed as other is seen in Table 3.


To evaluate variation in AI-predicted malignancy scores among revised benign histology categories, a one-way ANOVA was conducted. The analysis revealed no statistically significant differences between groups

($F=0.52$, $p=0.67$), suggesting that the AI system assigned comparable malignancy scores across distinct benign pathologies.

Discussion


This study demonstrates that the Breast AI system provides clinically meaningful support in identifying high-risk breast lesions, achieving a sensitivity of 67.2%, specificity of 79.4% and a PPV of 70.3% at a 51% malignancy threshold. These results underscore the system's potential utility as an adjunctive triage tool, particularly within resource-limited or high-volume clinical environments, by accurately flagging the majority of histologically confirmed malignant lesions for expedited referral.

Lateral discrepancy, as indicated in Table 2, warranted inclusion in this study as it may reflect real-world scanning challenges or biases introduced during model training. Prior research has shown anatomical and technical differences in image acquisition between breast sides – often influenced by operator handedness and transducer



AI, artificial intelligence.


FIGURE 2: Raincloud plot interpretation of artificial intelligence estimation (%) by histology result.



Note: Boxplots represent the median (horizontal line) and interquartile range (IQR, box). Whiskers extend to the most extreme data points within 1.5x IQR from the quartiles. Data points beyond this range are plotted individually as outliers (circles).

AI, artificial intelligence.

FIGURE 4: Artificial intelligence risk prediction score index for a benign diagnosis.



Note: Boxplots represent the median (horizontal line) and interquartile range (IQR, box). Whiskers extend to the most extreme data points within 1.5x IQR from the quartiles. Data points beyond this range are plotted individually as outliers (circles).

AI, artificial intelligence.

FIGURE 3: (a) Artificial intelligence (AI) prediction score versus (b) histology-confirmed malignancy rates by risk stratification.

angulation – which can affect lesion visibility and diagnostic accuracy.^{14,15} Moreover, AI models may be impacted by such confounding variables if laterality distribution is unbalanced in training data, potentially leading to skewed predictive performance.¹⁶ Reporting side-specific outcomes, therefore, offers valuable insight into generalisability and guides future refinements in AI training protocols to ensure equitable diagnostic performance.

The system's discriminatory capacity was further supported by the ROC curve analysis, which yielded an AUC of 0.76, indicating acceptable diagnostic performance. Using Youden's Index, the optimal malignancy risk threshold was initially identified at 51%, providing a pragmatic balance between sensitivity and specificity. Further recalibration using F1 score maximisation reaffirmed 51% as the optimal threshold, at which the model achieved its maximum F1 score of 65.7%. This suggests that a slightly lower threshold may better balance false positives and false negatives, depending on clinical context or risk tolerance. Area under the curve values in the range of 0.70–0.80 are generally considered clinically acceptable and support the feasibility of AI integration into diagnostic triage workflows, especially in settings with limited radiologist availability. A large, international study evaluating an AI system for breast cancer screening similarly concluded that AUCs within this range can justify clinical utility, particularly for triage or augmentation roles.¹⁷

A proposed three-tier risk stratification model – low-risk ($\leq 30\%$), intermediate-risk (31%–50%) and high-risk ($\geq 51\%$) – remains clinically relevant. The analysis confirmed that benign lesions predominantly clustered in the low-risk range, while most confirmed malignancies scored above the 51% threshold (Figure 5). However, the intermediate 'grey zone' (31%–50%) remains diagnostically ambiguous and may benefit from closer clinical correlation and short-interval follow-up rather than immediate intervention.

To further optimise the diagnostic balance between sensitivity and specificity, a performance curve was generated plotting F1 score, precision, and recall across AI-predicted malignancy thresholds (Figure 5). This analysis identified an optimal threshold of 51%, where the model achieved its maximum F1 score of 65.7%. The F1 score integrates both precision and recall, making it particularly relevant in oncologic triage where false negatives and false positives carry clinical consequences. This approach aligns with findings from Rodríguez-Ruiz et al.,¹⁸ who reported that F1-maximised thresholds yielded improved triage performance over fixed cut-offs in mammographic AI systems. Similarly, previous studies^{16,19} have emphasised the importance of threshold tuning in resource-constrained settings, where overly conservative thresholds may overburden referral systems, while overly liberal thresholds may compromise early detection. Compared to prior AI ultrasound studies reporting F1 scores between 68% and

75% in curated cohorts,^{16,19} the Breast AI system's current performance (F1 = 65.7%) reflects a reasonable balance under real-world conditions, particularly given the diverse case mix and imaging variability in this dataset. Continued refinement of threshold strategies – potentially incorporating cost-sensitive or patient-prioritised decision rules – may further enhance clinical integration.


While overall performance metrics are promising, sensitivity remains lower than the 85%–90% benchmarks commonly reported in curated datasets for AI-based ultrasound tools. Prior studies have documented higher sensitivity (76.9%–85.7%) and specificity (89.2%–96.1%) values in more controlled environments.^{17,20,21} This performance gap may reflect real-world variability in image quality, operator technique and case complexity, highlighting the need for ongoing algorithmic refinement and robust multicentre validation.^{22,23}

Importantly, the Breast AI system exhibited a relatively consistent pattern in its assessment of benign lesions. Subgroup analysis of common benign histologies – including fibroadenomas, fibroadipose tissue and benign breast parenchyma – showed mean AI-predicted malignancy scores ranging from 16.7% to 23.8%, with relatively narrow confidence intervals (e.g. fibroadenoma/fibroepithelial: 23.8%, 95% CI: 16.2%–31.4%). Slightly higher mean scores were noted for fat necrosis (27.1%, 95% CI: 17.8%–36.4%) and fibrocystic/inflammatory lesions (30.6%, 95% CI: 21.7%–39.4%), potentially because of sonographic overlap with malignant features. However, a one-way ANOVA

TABLE 3: Confidence interval rate for benign findings versus risk prediction score.

Benign category	Count	Mean	Standard	s.e.	95% CI	
					lower	upper
Fat necrosis	10	27.1	21.5	6.8	13.8	40.4
Fibroadenoma/ fibroepithelial	38	27.8	25.7	4.2	19.6	36.0
Inflammatory/fibrosis	22	29.7	25.7	5.5	18.9	40.4
Other benign entities	37	22.3	23.4	3.9	14.7	29.9

CI, confidence interval; s.e., standard error.



AI, artificial intelligence.


FIGURE 5: Precision-recall and F1 score threshold assessment.

showed no statistically significant differences among these groups ($F=0.52$, $p=0.67$), suggesting that the AI does not systematically overestimate malignancy risk within the benign spectrum. While this result should not be interpreted as evidence of equivalence, it offers preliminary support for the system's consistency across benign subtypes. Further multicentre studies are warranted to evaluate reproducibility and generalisability in varied imaging environments.^{24,25,26,27}

A focused review of false positives further supports this interpretation. Lesions such as fibroadenomas, chronic abscesses and postsurgical fibrosis were frequently flagged as high-risk because of complex echogenic patterns mimicking malignancy. These misclassifications likely reflect a conservative, safety-oriented bias rather than model error, consistent with findings from other AI-based diagnostic systems.¹⁸ In oncologic triage, such cautious classification may be clinically acceptable, provided it is balanced by a low rate of missed malignancies.^{28,29,30}

To enhance Breast AI's diagnostic performance and clinical utility, several refinements are recommended. Incorporating a dynamic threshold model guided by F1 maximisation or cost-sensitive adjustments could improve triage accuracy. Standardising image acquisition protocols and implementing quality control measures may help mitigate inter-operator variability. Additionally, integrating supplemental data such as Doppler flow, elastography, and relevant clinical metadata (e.g. hormonal status, genomic markers) may improve context-aware decision-making. Finally, prospective multicentre validation will be essential for confirming system generalisability and supporting regulatory approval.^{31,32}

Based on the diagnostic performance of the Breast AI system and threshold optimisation findings, we propose a three-tiered triage model to streamline clinical workflow and resource allocation, as illustrated in Figure 6:



AI, artificial intelligence; CBE, clinical breast examinations.

†, clinical breast examination (CBE) clinical signs and symptoms of breast cancer.

FIGURE 6: Risk-stratified clinical triage model for breast artificial intelligence.

1. *Low-risk lesions* (<30%) may be referred for routine follow-up or annual CBE, thereby reducing unnecessary imaging or biopsy.
2. *Intermediate-risk lesions* (30%–50%) represent a diagnostic grey zone and may benefit from short-term re-evaluation, adjunctive imaging, or clinical reassessment within 6–12 weeks.
3. *High-risk lesions* (≥51%), as supported by F1 score optimisation and malignancy clustering, should be prioritised for urgent referral to diagnostic centres.

This stratified triage pathway offers a pragmatic framework for efficient resource use while maintaining diagnostic vigilance. When deployed at the point of care, Breast AI may support timely clinical decision-making and help reduce delays in cancer diagnosis, particularly in underserved or high-throughput healthcare settings.³³

Conclusion

The Breast AI system demonstrates promising diagnostic performance, achieving a sensitivity of 67.2%, specificity of 79.4%, PPV of 70.3%, and an F1 score of 65.7% under real-world clinical conditions. These results highlight the system's potential utility as a triage and risk stratification tool in breast cancer care, particularly in settings where radiologist availability is limited and diagnostic backlogs are common. Its ability to identify higher-risk lesions and prioritise them for further investigation may help streamline referral workflows and reduce diagnostic delays.

Nonetheless, these findings must be interpreted within the context of the study's limitations, including a modest sample size, imaging heterogeneity and the absence of external or temporal validation. While subgroup analyses suggest consistency in benign lesion scoring and support a three-tier threshold-based triage approach, diagnostic ambiguity within the intermediate-risk range underscores the need for continued clinical oversight.

To enable broader clinical adoption, the Breast AI system will require further algorithmic refinement, standardisation of imaging protocols, and prospective validation across diverse clinical settings. These efforts are critical to ensuring reproducibility, regulatory compliance and clinical trustworthiness. If effectively integrated into routine workflows, Breast AI has the potential to enhance early detection and reduce delays in diagnosis – especially in underserved or high-throughput healthcare environments.

Acknowledgements

The article is based on a patented algorithm that has been formally registered under MedSol AI Pty Ltd and was previously presented in abstract form at BIGOSA, held in Durban on 29 November 2025. The abstract has since been developed into a full manuscript, which has been expanded and revised for journal publication. This republication is done with permission from the conference organisers.

Competing interests

The authors declare that they have no financial or personal relationships that may have inappropriately influenced them in writing this article.

Authors' contributions

K.M. was involved in the conceptualisation, methodology, formal analysis, investigation, writing of the original draft, visualisation, project administration, software development, validation, data curation, resource allocation, reviewing and editing of the article, supervision, and funding acquisition of this study. F.M. was involved in the conceptualisation, methodology, formal analysis, investigation, writing of the original draft, visualisation, project administration, validation, data curation, resources, reviewing and editing of the article, and supervision of this study. L.R. was involved in the methodology, writing of the original draft, validation, reviewing and editing of the article, and supervision of this study.

Funding information

This research received no specific grant from any funding agency in the public, commercial or not-for-profit sectors.

Data availability

The data that support the findings of this study are available on request from the corresponding author, K.M.

Disclaimer

The views and opinions expressed in this article are those of the authors and are the product of professional research. The article does not necessarily reflect the official policy or position of any affiliated institution, funder, agency or that of the publisher. The authors are responsible for this article's results, findings and content.

References

1. Swinny CA, Kagee A, Roomaney R. Delayed help-seeking for symptomatic breast cancer: Reasons for delay among participants receiving assessment at a public healthcare facility in South Africa. *S Afr J Psychol.* 2022;52(1):87–98. <https://doi.org/10.1177/0081246321992477>
2. Moodley J, Cairncross L, Naiker T, Constant D. From symptom discovery to assessment-women's pathways to breast cancer care: A cross-sectional study. *BMC Cancer.* 2018;18:1–11. <https://doi.org/10.1186/s12885-018-4219-7>
3. Dlamini Z, Molefi T, Khanyile R, et al. From incidence to intervention: A comprehensive look at breast cancer in South Africa. *Int J Environ Res Public Health.* 2023;20(21):7005. <https://doi.org/10.3390/ijerph20217005>
4. Sarmah N, Sibiya MN, Khoza TE. The sociocultural influences on breast cancer screening among rural African women in South Africa. *Int J Environ Res Public Health.* 2023;20(21):7005. <https://doi.org/10.3390/ijerph20217005>
5. Rayne S, Schnippel K, Firnhaber C, Wright K, Kruger D, Benn CA. Fear of treatments surpasses demographic and socioeconomic factors in affecting patients with breast cancer in urban South Africa. *J Glob Oncol.* 2017;3(2):125–134. <https://doi.org/10.1200/JGO.2015.002691>
6. Gonzaga MA. How accurate is ultrasound in evaluating palpable breast masses? *Pan Afr Med J.* 2010;7:1. Epub 2010 Sep 2. PMID: 21918690; PMCID: PMC3172638.
7. Youk JH, Kim EK, Kim MJ, Kwak JY, Son EJ. Performance of hand-held whole-breast ultrasound based on BI-RADS in women with mammographically negative dense breast. *Eur Radiol.* 2010;21(4):667–675. <https://doi.org/10.1007/s00330-010-1955-8>
8. Malherbe K. Revolutionizing breast cancer screening: Integrating artificial intelligence with clinical examination for targeted care in South Africa. *J Radiol Nurs.* 2025;44(2):195–202. <https://doi.org/10.1016/j.jradnu.2024.12.004>
9. Omidiji O, Campbell PC, Iruhe N, Atalabi OM, Toyobo O. Breast cancer screening in a resource poor country: Ultrasound versus mammography. *Ghana Med J.* 2017;51(1):6. <https://doi.org/10.4314/gmj.v51i1.2>
10. Mans P-A, Adeniyi OV. South African medical interns' perspectives on the use of point of care ultrasound. *S Afr Fam Pract.* 2023;65(1):a5772. <https://doi.org/10.4102/safp.v65i1.5772>
11. Sarmah N, Sibiya MN, Khoza TE. Barriers and enablers to breast cancer screening in rural South Africa. *Curationis.* 2024;47(1):a2583. <https://doi.org/10.4102/curationis.v47i1.2583>
12. Zheng D, He X, Jing J. Overview of artificial intelligence in breast cancer medical imaging. *J Clin Med.* 2023;12(2):419. <https://doi.org/10.3390/jcm12020419>
13. Malherbe K. A diagnostic algorithm for accurate detection of breast carcinoma on ultrasound [homepage on the Internet]. Doctoral dissertation, University of Pretoria; 2023 [cited 2021 Aug 18]. Available from: <https://repository.up.ac.za/items/2721960b-6eaa-4044-8a9e-e92b5df863d2>
14. D'Orsi CJ, Sickles EA, Mendelson EB, Morris EA. Breast imaging reporting and data system (BI-RADS) committee. ACR BI-RADS® atlas, breast imaging reporting and data system. Reston, VA: American College of Radiology; 2013.
15. Esteva A, Chou K, Yeung S, et al. Deep learning-enabled medical computer vision. *NPJ Digit Med.* 2019;2:10. <https://doi.org/10.1038/s41746-020-00376-2>
16. Zhu Y, Tan Y, Pang P. Deep learning of ultrasound images predicts histopathology of breast cancer. *J Ultrasound Med.* 2019;38(9):2509–2516. <https://doi.org/10.1002/jum.14942>
17. McKinney SM, Sieniek M, Godbole V, et al. International evaluation of an AI system for breast cancer screening. *Nature.* 2020;577(7788):89–94. <https://doi.org/10.1038/s41586-019-1799-6>
18. Rodríguez-Ruiz A, Lång K, Gubern-Merida A, et al. Stand-alone artificial intelligence for breast cancer detection in mammography: Comparison with 101 radiologists. *J Natl Cancer Inst.* 2019;111(9):916–922. <https://doi.org/10.1093/jnci/djy222>
19. Zhou LQ, Wu XL, Huang SY. Lymph node metastasis prediction from primary breast cancer US images using deep learning. *Radiology.* 2020;294(1):19–28. <https://doi.org/10.1148/radiol.2019190372>
20. Kwon H, Oh SH, Kim M-G, et al. Enhancing breast cancer detection through advanced AI-driven ultrasound technology: A comprehensive evaluation of visbus. *Diagnostics.* 2024;14(17):1867. <https://doi.org/10.3390/diagnostics14171867>
21. Koo C, Yang A, Welch CS, et al. Validating racial and ethnic non-bias of artificial intelligence decision support for diagnostic breast ultrasound evaluation. *J Med Imag.* 2023;10(06):061108. <https://doi.org/10.1117/1.JMI.10.6.061108>
22. Dickens C, Joffe M, Jacobson JS, et al. Stage at breast cancer diagnosis and distance from diagnostic hospital in a Periurban setting: A South African public hospital case series of over 1,000 women. *Int J Cancer.* 2014;135(9):2173–2182. <https://doi.org/10.1002/ijc.28861>
23. Tshabalala G, Blanchard C, Mmoledi K, et al. A qualitative study to explore healthcare providers' perspectives on barriers and enablers to early detection of breast and cervical cancers among women attending primary healthcare clinics in Johannesburg, South Africa. *PLoS Global Public Health.* 2023;3(5):e0001826. <https://doi.org/10.1371/journal.pgph.0001826>
24. Ha R, Mutasa S, Karcich J, et al. Predicting breast cancer molecular subtype with MRI-derived radiomics. *Radiology.* 2021;298(1):60–70.
25. Berg WA, Zhang Z, Lehner D, et al. Detection of breast cancer with addition of annual screening ultrasound or a single screening MRI to mammography in women with elevated breast cancer risk. *JAMA.* 2012;307(13):1394–1404. <https://doi.org/10.1001/jama.2012.388>
26. Wanders JOP, Veldhuis WB, Loo CE, et al. Volumetric breast density affects performance of digital screening mammography. *Breast Cancer Res Treat.* 2017;162(1):95–103. <https://doi.org/10.1007/s10549-016-4090-7>
27. Giuliano V, Giuliano C. Improved breast cancer detection using 3D automated breast ultrasound in women with dense breast tissue. *Clin Imaging.* 2013;37(3):480–486. <https://doi.org/10.1016/j.clinimag.2012.09.018>
28. Kim HE, Kim HH, Han BK, et al. Changes in cancer detection and false-positive recall in mammography using artificial intelligence: A retrospective, multireader study. *Lancet Digit Health.* 2020;2(3):e138–e148. [https://doi.org/10.1016/S2589-7500\(20\)30003-0](https://doi.org/10.1016/S2589-7500(20)30003-0)
29. Carty NJ, Carter C, Rubin C, Ravichandran D, Royle GT, Taylor I. Management of fibroadenoma of the breast. *Ann R Coll Surg Engl.* 1995 Mar;77(2):127–30. PMID: 7793802; PMCID: PMC2502143.
30. Youk JH, Gweon HM, Son EJ. Shear-wave elastography in breast ultrasonography: the state of the art. *Ultrasonography.* 2017 Oct;36(4):300–309. <https://doi.org/10.14366/usc.17024>. Epub 2017 Apr 5. PMID: 28513127; PMCID: PMC5621798.
31. Méndez CA, Pizzorni Ferrarese F, Summers P, Petralia G, Menegaz G. DCE-MRI and DWI Integration for Breast Lesions Assessment and Heterogeneity Quantification. *Int J Biomed Imaging.* 2012;2012:676808. <https://doi.org/10.1155/2012/676808>. Epub 2012 Nov 19. PMID: 23213317; PMCID: PMC3507154.
32. Skaane P, Young K, Skjennald A. Population-based mammography screening: Comparison of screen-film and full-field digital mammography with soft-copy reading—The Oslo I Study. *Radiology.* 2003;229(3):877–884. <https://doi.org/10.1148/radiol.2293021171>
33. Aliyu LD, Kurjak A, Wataganara T, et al. Ultrasound in Africa: What can really be done? *J Perinat Med.* 2016;44(2):119–123. <https://doi.org/10.1515/jpm-2015-0224>